

Topweigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe, Carlisle, Cumbria. CA6 4BH. UK



Tel: +44(0)1228 672400, Fax: +44(0)7092 884848 Mobile: +44(0)7808 308689, Email: ajj@Scalemaker.co.uk

Calibration Certificate

Customer				Customer Order No.		
Dundas Chemical Company Ltd				Date of Last Test	15 May 2021	
				Date of Test	5 September 2021	
				Next Test Due	5 March 2022	
Site Address :-				Next Test Due within 30 days of above Due date.		
Oakbank Services				Machine	Weighbridge	
Mosspark				Туре		
Annan Road				Manufacturer	Topweigh	
Dumfries				Capacity :- Max	50,000 Kg	
DG1 4PH				Min	400 Kg	
				e.	20 Kg	
Calibration Certificate No.		TW	1021186	Maximum Tolerance	3e	
				Serial No.	TW-D-15-50-01	
	Reading	Weights	Reading			
Applied		Applied		Class Observations & Comm	((III)	
				within Inspection tolerand weighing Instruments reground to be CORRECT Weights Traceable to Note that the Cortificate No.	Traceable to National Standards te No. TW 10783	

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :- Print Name :- A J Jeffrey
Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.